

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 03 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130191</u> ✓
Name <u>BRANDON LAWRENSEN</u>		ID# <u>6609</u> Date <u>11/25/09</u> ✓
A	Agency <u>JUNEAU POLICE DEPT</u>	Phone # <u>586-0600</u>
Instrument Location <u>JUNEAU POLICE DEPT</u> ✓		
Alco S/N <u>X124634</u> ✓		Target Value <u>0.088*</u> ✓ ^{008/22/10} High Pressure <u>10000</u>
B	Alco Test Values	
	<u>.093</u> ✓ 1 st Alco	<u>.094</u> ✓ 2 nd Alco
Signature <u>[Signature]</u> ✓		<u>008</u> 1/22/10
(OVER)		

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 28th day of January, 2009.
[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

NOVEMBER 25, 2009

OPERATOR'S NAME: *COS 1/22/10*
LAWRENSEN/BRANDON/J
OPERATOR'S NUMBER: 6609
SUBJECT'S LAST NAME:

VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 0123456
DEPT/AGENCY: JNU1
CASE/REPORT: 0123456
TEST TYPE: V
ALCO TARGET VALUE: ~~0.086~~ * *0.088 COS 1/22/10*
ALCO S/N: X124634

--- BREATH ANALYSIS ---

.086 ADJUSTED FOR 29.33 in		
ALCO TARGET	.084	18:59
BLANK TEST	.000	19:00
INTERNAL STANDARD	VERIFIED	19:00
ALCO TV 29.33 in	.093	19:00
BLANK TEST	.000	19:01
SUBJECT SAMPLE <i>COS 1/22/10</i>	.000	19:02
BLANK TEST	.000	19:02
ALCO TV 29.33 in	.094	19:03
BLANK TEST	.000	19:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

NOVEMBER 25, 2009
TIME 18:39

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS:	
SAMPLE CHAMBER:	50c
BREATH TUBE:	42c
BAROMETER:	29.34 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~

* Target Value corrected at JCDL.
COS 1/22/10